

---

Poster

## Fine tuning and interference study in the determination of heavy metals in fishery products by ICP-MS technique



María Victoria Sánchez (1,\*), María de la Menta Ballesteros (1), Raquel Rojas(2)

(1)Departamento de Biología Molecular e Ingeniería Bioquímica/Universidad Pablo de Olavide, Ctra. de Utrera Km 1, 41013, Sevilla.

(2)Departamento Físico Químico/Laboratorios Vital, C. Imprenta nº28, P.I. La Negrilla, 41016, Sevilla.

**Keywords:** ICP-MS; heavy metals; fishery products; internal standard; cadmium

---

### ABSTRACT

Fish is one of the main protein sources for human diet which is usually affected by the pollution of aquatic ecosystem due to bioaccumulation of heavy metals, among others. For this reason, these pollutants are included in the regulation of the European Community Commission which set the maximum levels for certain contaminants in foodstuff. In order to quantify heavy metals, Inductively Coupled Plasma Mass Spectrometry (ICP-MS) is a great analytical technique due to its high accuracy and precision in detecting metals at low concentrations. The aim of this project is to carry out a fine tuning procedure and interference elimination studies during the determination of mercury (Hg), cadmium (Cd) and lead (Pb) in fishery products using ICP-MS. Linearity was studied by reduction of the curve range in order to increase the accuracy of the method. Moreover, selection of the optimum internal standard in order to correct matrix effect during the standardization process was evaluated. Residuals and percentage of internal standard recovery are the quality controls established to determine if the selected standard (Rh-103) was adequate. Finally, results indicate that Cd-111 was the isotope more suitable in order to avoid the spectral interferences.

### REFERENCES

- Djedjibegovic, J., Larssen, T., Skrbo, A., Marjanović, A., & Sober, M. (2012). Contents of cadmium, copper, mercury and lead in fish from the Neretva river (Bosnia and Herzegovina) determined by inductively coupled plasma mass spectrometry (ICP-MS). *Food Chemistry*, 131(2), 469-476.
- Guo, W., Hu, S., Zhao, J., Jin, S., Liu, W., & Zhang, H. (2011). Removal of spectral interferences and accuracy monitoring of trace cadmium in feeds by dynamic reaction cell inductively coupled plasma mass spectrometry. *Microchemical Journal*, 97(2), 154-159.
- Nardi, E. P., Evangelista, F. S., Tormen, L., Saint, T. D., Curtius, A. J., de Souza, S. S., & Barbosa, F. (2009). The use of inductively coupled plasma mass spectrometry (ICP-MS) for the determination of toxic and essential elements in different types of food samples. *Food Chemistry*, 112(3), 727-732.